


## Measurement Conditions

**Measurement setup: CAS 140B/CT(S)/D, CAS 120, MAS 40 - 2365142E**

|   |                            |
|---|----------------------------|
| Integration time: 55 ms                         | Averages: 20               |
| Density filter: none                            | Serial number: 236514212   |
| Signal level: 45 % 13849 Counts                 | CCD temperature: -10,18 °C |
| Analysis range 300 - 1000 nm                    | Last dark current: 0,0 min |
| Illuminant ref.: Illuminant D65 (0,3138;0,3310) |                            |
| Skiplevel:                                      | Observer: 10 °             |

## Spectrum PFCL B



## Results

|                    |           |                           |     |        |
|--------------------|-----------|---------------------------|-----|--------|
| Y Tau%             | 36,52     | <b>Color coordinates:</b> | x:  | 0,2002 |
| Peak WL:           | 409,73 nm |                           | y:  | 0,2292 |
| Peak Transmission: | 0,90      |                           | u:  | 0,1497 |
| Width 50%:         | 0,00 nm   |                           | v:  | 0,2570 |
| Width 10%:         | 0,00 nm   |                           | u': | 0,1497 |
| Width 1%:          | 0,00 nm   |                           | v': | 0,3856 |

### L\*a\*b\*

L\*: 66,912  
a\*: -9,6554  
b\*: -46,286  
Ca\*b\*: -47,283  
Ha\*b\*: -258,22

### L\*u\*v\*

L\*: 66,912  
u\*: -41,929  
v\*: -73,046  
Cu\*v\*: -66,912  
Hu\*v\*: -9,6554

Comment:

File name: C:\Users\Laptop-01\Desktop\Misure\_Filtri\QB1.ISD

Date: 25/10/2019 10:52:18